

Analysenzertifikat

Certificate of Analysis

SERION KBR Antigen / CFT Antigen INFLUENZA A VIRUS

Best.-Nr. / Order No.	1112
Chargen-Nr. / Lot No.	KL0028
IFU-Version / IFU version	21-19/07-1
Verfallsdatum / Expiry Date	2030-03-31
Antigen / Antigen	Influenza A Virus (Wisconsin)
Kultursystem / Culture System	bebrütete Hühnereier incubated hen's eggs
Präparation / Preparation	konzentrierte Virus Core-Partikel concentrated virus core particles
Inaktivierung / Inactivation	Behandlung mit Detergens treatment with detergent
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:32



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de