

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	EK0163
Verfallsdatum / Expiry Date	2021-12-31
IFU-Version / IFU-Version	NEW! 131-14

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECK0349	0,94	> 0,47
Cut off-Serum / Cut off serum	ECK0351	0,36	0,18 - 0,72

Negativkontrolle / Negative Control	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECK0348
Teststreifen / Strips	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECK0350
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJK006

10.01.2020

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de