

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	EK0042
Verfallsdatum / Expiry Date	2021-04
IFU-Version / IFU-Version	NEW! 131-14

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	SDK.DZ	1,17	> 0,59
Cut off-Serum / Cut off serum	SDK.EG	0,59	0,30-1,18

Negativkontrolle / Negative Control Chargen-Nr. / Lot No.	COXIELLA BURNETII PHASE 1 IgA SDK.DX
Teststreifen / Strips Chargen-Nr. / Lot No.	COXIELLA BURNETII PHASE 1 IgA SDK.DS
Konjugat / Conjugate Chargen-Nr. / Lot No.	IgA ++ KJK001

06.06.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de