

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 1 IgG

Best.-Nr. / Order No.	ESR1311G
Chargen-Nr. / Lot No.	SCK.ET
Verfallsdatum / Expiry Date	2021-02
IFU-Version / IFU-Version	NEW! 131-14

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	SBK.GE	0,92	> 0,46
Cut off-Serum / Cut off serum	SBK.GF	0,45	0,22 – 0,90

Negativkontrolle / Negative Control	COXIELLA BURNETII PHASE 1 IgG
Chargen-Nr. / Lot No.	SBK.GD
Teststreifen / Strips	COXIELLA BURNETII PHASE 1 IgG
Chargen-Nr. / Lot No.	VBI.AA
Konjugat / Conjugate	IgG +++
Chargen-Nr. / Lot No.	SCK.CP

28.03.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de