

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	SBK.BL
Verfallsdatum / Expiry Date	2020-11
IFU-Version / IFU-Version	NEW! 131-14

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	SMI.DE	1,02	> 0,51
Cut off-Serum / Cut off serum	SMI.DF	0,64	0,32 – 1,28

Negativkontrolle / Negative Control	COXIELLA BURNETII PHASE 1 IgA
Chargen-Nr. / Lot No.	SMI.CX
Teststreifen / Strips	COXIELLA BURNETII PHASE 1 IgA
Chargen-Nr. / Lot No.	VLH.AE
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	SAK.EM

12.02.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de