

# Analysenzertifikat Certificate of Analysis

**SERION KBR Kontrollantigen / CFT Control Antigen  
COXSACKIEVIRUS A9**

Best.-Nr. / Order No.	9260
Chargen-Nr. / Lot No.	KL0060
IFU-Version / IFU version	21-19/07-1
Verfallsdatum / Expiry Date	2030-07-31
Kultursystem / Culture System	LLC-MK2 Zellen / cells
Präparation / Preparation	Alkalisch extrahierte Zellen (konzentriert und gereinigt) Cells extracted with alkaline buffer (concentrated and purified)
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:4



Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de