

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* West Nile Virus IgM

Best.-Nr. / Order No.	ESR141M
Chargen-Nr. / Lot No.	EL0182
Verfallsdatum / Expiry Date	2022-11-30
IFU-Version / IFU-Version	<b>NEW!</b> 141-4

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECL0472	0,87	> 0,44
Cut off-Serum / Cut off serum	ECL0473	0,19	0,10 – 0,38

Negativkontrolle / Negative Control	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECL0459
Teststreifen / Strips	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECL0467
Konjugat / Conjugate	IgM +++
Chargen-Nr. / Lot No.	KJL012

30.11.2020

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de