

# Analysezertifikat Certificate of Analysis

## SERION KBR Negativkontrolle / CFT Negative Control COXIELLA BURNETII PHASE 1

Best.-Nr. / Order No.	4227
Chargen-Nr. / Lot No.	KL0071
IFU-Version / IFU version	21-19/07-1
Verfallsdatum / Expiry Date	2030-05-31
Material / Material	Fötiales Kälberserum / Fetal calf serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	< 1:10



Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de