

**Qualitätskontrollzertifikat**  
**Quality Control Certificate**



**SERION ELISA agile**  
**SARS-CoV-2 IgM**

Best.-Nr. / Order No.	ESR400M
Chargen-Nr. / Lot No.	AL046
Verfallsdatum / Expiry Date	2022-11-30
IFU-Version / IFU-Version	a400M-1

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	AL043	0,787	> 0,394
Cut off-Serum / Cut off serum	AL044	0,322	0,161-0,644

Negativkontrolle / Negative Control Chargen-Nr. / Lot No.	AL042
Teststreifen / Strips Chargen-Nr. / Lot No.	AL041
Konjugat / Conjugate Chargen-Nr. / Lot No.	AL045
RF Absorbent (Z23) / RF Absorbent (Z23) Chargen-Nr. / Lot No.	AL037

14.12.2020

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de

FB-IV-312A