

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgG

Best.-Nr. / Order No.	ESR1311G
Chargen-Nr. / Lot No.	EM0087
Verfallsdatum / Expiry Date	2023-03-31
IFU-Version / IFU-Version	NEW! 131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECM0129	0,75	> 0,38
Cut off-Serum / Cut off serum	ECM0130	0,37	0,19 – 0,74

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECM0128
Teststreifen / Strips	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECM0103
Konjugat / Conjugate	IgG +++
Chargen-Nr. / Lot No.	KJL015

18.03.2021

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de