

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	EL0103
Verfallsdatum / Expiry Date	2022-04-30
IFU-Version / IFU-Version	NEW! 131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECL0237	1,11	> 0,56
Cut off-Serum / Cut off serum	ECL0238	0,59	0,30 – 1,18

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECL0236
Teststreifen / Strips	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	SDK.DS
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	KJL001

03.08.2020

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de