

**Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for
Revelation™ DSX / DS-Matrix™****4PS- Formel / 4PS-formula**
$$\exp(3.760 - \ln(2.078 / (\text{Sample}^{0.730} / S - 0.009) - 1) / 1.031)$$
**Gültigkeitsbereich / Validity Range**
$$0.365 \leq S1 \leq 1.241$$
**If OD Sample < Parameter A**
$$\text{if } Ti < (0.009 * (S1 / 0.730)) \text{ then } Ti = (0.009 + 0.001) * (S1 / 0.730)$$
**If OD Sample > Parameter D**
$$\text{if } Ti > (2.087 * (S1 / 0.730)) \text{ then } Ti = (2.087 - 0.001) * (S1 / 0.730)$$
**If OD Negative control < Parameter A**
$$\text{if } NC1 < (0.009 * (S1 / 0.730)) \text{ then } NCi = (0.009 + 0.001) * (S1 / 0.730)$$
