

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	EM0198
Verfallsdatum / Expiry Date	2023-09-30
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECM0398	1,02	> 0,51
Cut off-Serum / Cut off serum	ECM0399	0,68	0,34 – 1,36

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECM0397
Teststreifen / Strips	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECM0396
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	KJM024

27.09.2021

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de