

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgG

Best.-Nr. / Order No.	ESR1311G
Chargen-Nr. / Lot No.	EM0219
Verfallsdatum / Expiry Date	2023-10-31
IFU-Version / IFU-Version	NEW! 131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECM0448	0,47	> 0,24
Cut off-Serum / Cut off serum	ECM0449	0,32	0,16 – 0,64

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECM0447
Teststreifen / Strips	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECM0103
Konjugat / Conjugate	IgG +++
Chargen-Nr. / Lot No.	KJM023

04.11.2021

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de