

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* West Nile Virus IgM

Best.-Nr. / Order No.	ESR141M
Chargen-Nr. / Lot No.	EN0046
Verfallsdatum / Expiry Date	2024-02-29
IFU-Version / IFU-Version	141-4

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECN0073	0,91	> 0,46
Cut off-Serum / Cut off serum	ECN0078	0,24	0,12 – 0,48

Negativkontrolle / Negative Control	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECN0072
Teststreifen / Strips	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECN0084
Konjugat / Conjugate	IgM +++
Chargen-Nr. / Lot No.	KJM026

28.02.2022

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de