

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	EN0076
Verfallsdatum / Expiry Date	2024-03-31
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECN0122	1,04	> 0,52
Cut off-Serum / Cut off serum	ECN0123	0,61	0,31 – 1,22

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECN0121
Teststreifen / Strips	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECN0150
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	KJN029

12.04.2022

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de