

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Respiratory Syncytial Virus IgM

Best.-Nr. / Order No.	ESR113M
Chargen-Nr. / Lot No.	EM0236
Verfallsdatum / Expiry Date	2023-11-30
IFU-Version / IFU-Version	113-16

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECM0493	0,93	> 0,47
Cut off-Serum / Cut off serum	ECM0508	0,29	0,15 – 0,58

Negativkontrolle / Negative Control	Respiratory Syncytial Virus IgM
Chargen-Nr. / Lot No.	ECM0492
Teststreifen / Strips	Respiratory Syncytial Virus IgM
Chargen-Nr. / Lot No.	ECM0362
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJM025

19.11.2021

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de