



## Analysezertifikat Certificate of Analysis

### SERION ELISA *classic* Avidity Reagent CYTOMEGALOVIRUS

Best.-Nr. / Order No.	B109AVID
Chargen-Nr. / Lot No.	ECN0241
IFU-Version / IFU version	7-19/07-1
Herstelldatum / Manufacturing date	2022-07
Verfallsdatum / Expiry Date	2032-05-31
Verwendung mit / For use with SERION ELISA <i>classic</i> Cytomegalovirus IgG	EM0049
Verwendung mit Standardserum / For use with standard serum	ECM0086
Standard Zielwert (SERION Aviditätsindex) / Standard Target Vange (SERION Avidity Index)	97 %
Standard Zielbereich (SERION Aviditätsindex) / Standard Target Range (SERION Avidity Index)	≥ 77,6 % und / and ≤ 116,4 %
X Faktor / X Factor	0,162
Y Faktor / Y Factor	1,203
Rekonstitution / Reconstitution	1,5 ml
Versions-Nr. / Version No.	2022 / 07 - 2



Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [info@virion-serion.de](mailto:info@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)

FB-IV-165 D