

# Analysenzertifikat

## Certificate of Analysis

### SERION KBR Positivkontrolle / CFT Positive Control ROTAVIRUS

Best.-Nr. / Order No.	3193
Chargen-Nr. / Lot No.	KN0012
IFU-Version / IFU version	22-20/02-1
Verfallsdatum / Expiry Date	2032-08-31
Material / Material	Humanserum / human serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	1:80



Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: info@virion-serion.de  
Internet: www.virion-serion.de