

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* Adenovirus IgM

Best.-Nr. / Order No.	ESR128M
Chargen-Nr. / Lot No.	EN0262
Verfallsdatum / Expiry Date	2024-12-31
IFU-Version / IFU-Version	128-16

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECN0580	0,75	> 0,38
Cut off-Serum / Cut off serum	ECN0581	0,28	0,14 – 0,56

Negativkontrolle / Negative Control	Adenovirus IgM
Chargen-Nr. / Lot No.	ECN0579
Teststreifen / Strips	Adenovirus IgM
Chargen-Nr. / Lot No.	ECN0586
Konjugat / Conjugate	IgM ++
Chargen-Nr. / Lot No.	KJN043

16.01.2023

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de