

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	E00030
Verfallsdatum / Expiry Date	2025-02-28
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0064	0,81	> 0,41
Cut off-Serum / Cut off serum	ECO0069	0,44	0,22 – 0,88

Negativkontrolle / Negative Control	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECO0063
Teststreifen / Strips	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECO0070
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJN046

24.02.2023

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de