

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* Coxiella burnetii Phase 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	E00077
Verfallsdatum / Expiry Date	2025-03-31
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0129	1,15	> 0,57
Cut off-Serum / Cut off serum	ECO0130	0,63	0,32 – 1,26

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECO0128
Teststreifen / Strips	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECN0600
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	KJ0053++

05.05.2023

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de