

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Adenovirus IgM

Best.-Nr. / Order No.	ESR128M
Chargen-Nr. / Lot No.	E00129
Verfallsdatum / Expiry Date	2025-06-30
IFU-Version / IFU-Version	128-16

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0282	0,82	> 0,41
Cut off-Serum / Cut off serum	ECO0286	0,36	0,18 – 0,72

Negativkontrolle / Negative Control	Adenovirus IgM
Chargen-Nr. / Lot No.	ECO0281
Teststreifen / Strips	Adenovirus IgM
Chargen-Nr. / Lot No.	ECO0273
Konjugat / Conjugate	IgM ++
Chargen-Nr. / Lot No.	KJ0057

20.07.2023

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de