

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* West Nile Virus IgM

Best.-Nr. / Order No.	ESR141M
Chargen-Nr. / Lot No.	E00147
Verfallsdatum / Expiry Date	2025-07-31
IFU-Version / IFU-Version	141-4

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0316	0,87	> 0,44
Cut off-Serum / Cut off serum	ECO0317	0,27	0,14 – 0,54

Negativkontrolle / Negative Control	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECO0315
Teststreifen / Strips	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECO0322
Konjugat / Conjugate	IgM +++
Chargen-Nr. / Lot No.	KJO050

11.08.2023

Prüfdatum / Date of control

---

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de