

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgG

Best.-Nr. / Order No.	ESR1311G
Chargen-Nr. / Lot No.	E00215
Verfallsdatum / Expiry Date	2025-11-30
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0472	0,92	> 0,46
Cut off-Serum / Cut off serum	ECO0473	0,54	0,27 – 1,08

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECO0471
Teststreifen / Strips	Coxiella burnetii Phase 1 IgG
Chargen-Nr. / Lot No.	ECO0478
Konjugat / Conjugate	IgG +++
Chargen-Nr. / Lot No.	KJ0064

02.01.2024

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de