

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Respiratory Syncytial Virus IgM

Best.-Nr. / Order No.	ESR113M
Chargen-Nr. / Lot No.	EP0005
Verfallsdatum / Expiry Date	2025-10-31
IFU-Version / IFU-Version	113-16

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECO0426	0,63	> 0,32
Cut off-Serum / Cut off serum	ECO0427	0,21	0,11 – 0,42

Negativkontrolle / Negative Control	Respiratory Syncytial Virus IgM
Chargen-Nr. / Lot No.	ECO0425
Teststreifen / Strips	Respiratory Syncytial Virus IgM
Chargen-Nr. / Lot No.	ECO0430
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJO065

17.01.2024

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de