

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* West Nile Virus IgM

Best.-Nr. / Order No.	ESR141M
Chargen-Nr. / Lot No.	EP0073
Verfallsdatum / Expiry Date	2026-02-28
IFU-Version / IFU-Version	141-4

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECP0049	0,97	> 0,48
Cut off-Serum / Cut off serum	ECP0050	0,27	0,14 – 0,54

Negativkontrolle / Negative Control	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECP0048
Teststreifen / Strips	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECP0053
Konjugat / Conjugate	IgM +++
Chargen-Nr. / Lot No.	KJP067

11.03.2024

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de