



Analysenzertifikat Certificate of Analysis

SERION ELISA *classic* Avidity Reagent TOXOPLASMA GONDII

Best.-Nr. / Order No.	B110AVID
Chargen-Nr. / Lot No.	ECP0065
IFU-Version / IFU version	11-22/10-1
Herstelldatum / Manufacturing date	2024-02
Verfallsdatum / Expiry Date	2034-02-28
Verwendung mit / For use with SERION ELISA <i>classic</i> Toxoplasma gondii IgG	EP0135
Verwendung mit Standardserum / For use with standard serum	ECP0170
Standard Zielwert (SERION Aviditätsindex) / Standard Target Value (SERION Avidity Index)	64 %
Standard Zielbereich (SERION Aviditätsindex) / Standard Target Range (SERION Avidity Index)	≥ 51,2 % und / and ≤ 76,8 %
X Faktor / X Factor	0,184
Y Faktor / Y Factor	1,054
Rekonstitution / Reconstitution	2,0 ml
Versions-Nr. / Version No.	2024 / 06 - 2



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: info@virion-serion.de
Internet: www.virion-serion.de