

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	EP0091
Verfallsdatum / Expiry Date	2026-03-31
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECP0087	0,82	> 0,41
Cut off-Serum / Cut off serum	ECP0088	0,43	0,22 – 0,86

Negativkontrolle / Negative Control Chargen-Nr. / Lot No.	COXIELLA BURNETII PHASE 2 IgM ECP0086
Teststreifen / Strips Chargen-Nr. / Lot No.	COXIELLA BURNETII PHASE 2 IgM ECP0082
Konjugat / Conjugate Chargen-Nr. / Lot No.	IgM + KJP071

16.04.2024

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de