

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* COXIELLA BURNETII PHASE 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	EP0199
Verfallsdatum / Expiry Date	2026-07-31
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECP0308	0,77	> 0,39
Cut off-Serum / Cut off serum	ECP0309	0,31	0,16 – 0,62

Negativkontrolle / Negative Control	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECP0307
Teststreifen / Strips	COXIELLA BURNETII PHASE 2 IgM
Chargen-Nr. / Lot No.	ECP0317
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJP071

07.08.2024

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de