

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* Coxiella burnetii Phase 1 IgA

Best.-Nr. / Order No.	ESR1311A
Chargen-Nr. / Lot No.	EP0284
Verfallsdatum / Expiry Date	2026-12-31
IFU-Version / IFU-Version	131-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECP0525	1,18	> 0,59
Cut off-Serum / Cut off serum	ECP0526	0,51	0,26 – 1,02

Negativkontrolle / Negative Control	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECP0524
Teststreifen / Strips	Coxiella burnetii Phase 1 IgA
Chargen-Nr. / Lot No.	ECP0518
Konjugat / Conjugate	IgA ++
Chargen-Nr. / Lot No.	KJP078++

03.01.2025

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de