



**Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for  
Revelation™ DSX / DS-Matrix™****4PS- Formel / 4PS-formula**
$$\exp(4.322 - \ln(2.413 / (\text{Sample}^{0.760} / S + 0.004) - 1) / 0.960)$$
**Gültigkeitsbereich / Validity Range**
$$0.380 \leq S1 \leq 1.292$$
**If OD Sample < Parameter A**
$$\text{if } Ti < (-0.004 * (S1 / 0.760)) \text{ then } Ti = (-0.004 + 0.001) * (S1 / 0.760)$$
**If OD Sample > Parameter D**
$$\text{if } Ti > (2.409 * (S1 / 0.760)) \text{ then } Ti = (2.409 - 0.001) * (S1 / 0.760)$$
**If OD Negative control < Parameter A**
$$\text{if } NC1 < (-0.004 * (S1 / 0.760)) \text{ then } NCi = (-0.004 + 0.001) * (S1 / 0.760)$$
