

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	EQ0068
Verfallsdatum / Expiry Date	2027-03-31
IFU-Version / IFU-Version	NEW! 131-16

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECQ0150	0,89	> 0,45
Cut off-Serum / Cut off serum	ECQ0151	0,40	0,20 - 0,80

Negativkontrolle / Negative Control	Coxiella burnetii Phase 2 IgM
Chargen-Nr. / Lot No.	ECQ0149
Teststreifen / Strips	Coxiella burnetii Phase 2 IgM
Chargen-Nr. / Lot No.	ECQ0146
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJP080

01.04.2025

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de