

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* West Nile Virus IgM

Best.-Nr. / Order No.	ESR141M
Chargen-Nr. / Lot No.	EQ0070
Verfallsdatum / Expiry Date	2027-03-31
IFU-Version / IFU-Version	NEW! 141-5

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECQ0133	0,96	> 0,48
Cut off-Serum / Cut off serum	ECQ0134	0,25	0,13 – 0,50

Negativkontrolle / Negative Control	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECQ0132
Teststreifen / Strips	West Nile Virus IgM
Chargen-Nr. / Lot No.	ECQ0128
Konjugat / Conjugate	IgM +++
Chargen-Nr. / Lot No.	KJQ085

01.04.2025

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de