

# Qualitätskontrollzertifikat

## Quality Control Certificate

### SERION ELISA *classic* Coxiella burnetii Phase 2 IgM

Best.-Nr. / Order No.	ESR1312M
Chargen-Nr. / Lot No.	ER0011
Verfallsdatum / Expiry Date	2027-03-31
IFU-Version / IFU-Version	<b>NEW!</b> 131-17

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	ECQ0150	0,98	> 0,49
Cut off-Serum / Cut off serum	ECR0021	0,42	0,21 - 0,84

Negativkontrolle / Negative Control	Coxiella burnetii Phase 2 IgM
Chargen-Nr. / Lot No.	ECQ0149
Teststreifen / Strips	Coxiella burnetii Phase 2 IgM
Chargen-Nr. / Lot No.	ECQ0146
Konjugat / Conjugate	IgM +
Chargen-Nr. / Lot No.	KJP080

22.01.2026

Prüfdatum / Date of control

Hersteller / Manufacturer:  
 Institut Virion\Serion GmbH  
 Friedrich-Bergius-Ring 19  
 D-97076 Würzburg  
 Germany  
 Tel. +49 931 3045 0  
 Fax +49 931 3045 100  
 E-Mail: info@virion-serion.de  
 Internet: www.virion-serion.de