

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* Coxiella burnetii Phase 1 IgG

| | |
|-----------------------------|--------------------|
| Best.-Nr. / Order No. | ESR1311G |
| Chargen-Nr. / Lot No. | ER0013 |
| Verfallsdatum / Expiry Date | 2028-01-31 |
| IFU-Version / IFU-Version | NEW! 131-17 |

| Kontrollseren / Control sera | Chargen-Nr. / Lot No | Gemessener Wert / Measured value OD 405 nm | Gültigkeitsbereich / Validity range |
|--|-------------------------|--|--|
| Positivkontrolle / Positive control | ECR0023 | 1,21 | > 0,61 |
| Cut off-Serum / Cut off serum | ECR0024 | 0,48 | 0,24 – 0,96 |

| | |
|-------------------------------------|-------------------------------|
| Negativkontrolle / Negative Control | Coxiella burnetii Phase 1 IgG |
| Chargen-Nr. / Lot No. | ECR0022 |
| Teststreifen / Strips | Coxiella burnetii Phase 1 IgG |
| Chargen-Nr. / Lot No. | ECR0036 |
| Konjugat / Conjugate | IgG +++ |
| Chargen-Nr. / Lot No. | KJQ091 |

28.01.2026

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de