


**Zusätzliche Barcodes mit Formeln für / Additional Barcodes with formulas for
Revelation™ DSX / DS-Matrix™****4PS- Formel / 4PS-formula**
$$\exp(4.965 - \ln(4.219 / (\text{Sample}^{0.770} / S + 0.115) - 1) / 0.557)$$
**Gültigkeitsbereich / Validity Range**
$$0.385 \leq S1 \leq 1.309$$
**If OD Sample < Parameter A**
$$\text{if } Ti < (-0.115 * (S1 / 0.770)) \text{ then } Ti = (-0.115 + 0.001) * (S1 / 0.770)$$
**If OD Sample > Parameter D**
$$\text{if } Ti > (4.104 * (S1 / 0.770)) \text{ then } Ti = (4.104 - 0.001) * (S1 / 0.770)$$
**If OD Negative control < Parameter A**
$$\text{if } NC1 < (-0.115 * (S1 / 0.770)) \text{ then } NCi = (-0.115 + 0.001) * (S1 / 0.770)$$
