



Analysenzertifikat Certificate of Analysis

SERION ELISA *classic* Avidity Reagent RUBELLA VIRUS

Best.-Nr. / Order No.	B129AVID
Chargen-Nr. / Lot No.	ECQ0403
IFU-Version / IFU version	12-26/03-1
Herstelldatum / Manufacturing date	2025-08
Verfallsdatum / Expiry Date	2035-08-31
Verwendung mit / For use with SERION ELISA <i>classic</i> Röteln/Rubella Virus IgG	EQ0165
Verwendung mit Standardserum / For use with standard serum	ECQ0325
Standard Zielwert (SERION Aviditätsindex) / Standard Target Value (SERION Avidity Index)	83 %
Standard Zielbereich (SERION Aviditätsindex) / Standard Target Range (SERION Avidity Index)	≥ 66,4 % und / and ≤ 99,6 %
X Faktor / X Factor	0,377
Y Faktor / Y Factor	1,960
Rekonstitution / Reconstitution	2,0 ml
Versions-Nr. / Version No.	2026 / 03 - 3



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: info@virion-serion.de
Internet: www.virion-serion.de

FB-IV-165 H