

Qualitätskontrollzertifikat

Quality Control Certificate

SERION ELISA *classic* ADENOVIRUS IgM

Best.-Nr. / Order No.	ESR128M
Chargen-Nr. / Lot No.	EK0036
Verfallsdatum / Expiry Date	2021-01
IFU-Version / IFU-Version	NEW! 128-15

Kontrollseren / Control sera	Chargen-Nr. / Lot No	Gemessener Wert / Measured value OD 405 nm	Gültigkeitsbereich / Validity range
Positivkontrolle / Positive control	SAK.DP	0,95	> 0,48
Cut off-Serum / Cut off serum	SAK.DQ	0,27	0,14 – 0,54

Negativkontrolle / Negative Control	ADENOVIRUS IgM
Chargen-Nr. / Lot No.	SAK.DO
Teststreifen / Strips	ADENOVIRUS IgM
Chargen-Nr. / Lot No.	SDK.DY
Konjugat / Conjugate	IgM ++
Chargen-Nr. / Lot No.	SKI.EL

05.06.2019

Prüfdatum / Date of control

Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de