



# Analysenzertifikat

## Certificate of Analysis

virion\serion

### SERION ELISA *classic* Aviditätsreagenz / Avidity Reagent RÖTELN VIRUS / RUBELLA VIRUS

Best.-Nr. / Order No.	B129AVID
Chargen-Nr. / Lot No.	SID.CI
Verfallsdatum / Expiry Date	2023-09
Verwendung mit / For use with SERION ELISA <i>classic</i> Röteln Virus / Rubella Virus IgG	SKD.DH
Verwendung mit Standardserum / For use with standard serum	SKD.DD
Standard Zielwert (SERION Aviditätsindex) / Standard Target Value (SERION Avidity Index)	Zielwert 67 %
Standard Zielbereich (SERION Aviditätsindex) / Standard Target Range (SERION Avidity Index)	$\geq 54$ % und / and $\leq 81$ %
X Faktor / X factor	0,420
Y Faktor / Y factor	1,891
Rekonstitution / Reconstitution	2,0 ml Aqua dest.
Versions-Nr. / Version No.	2014 / 08-1

Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)