



Analysenzertifikat

Certificate of Analysis

virion\serion

SERION ELISA *classic* Aviditätsreagenz / Avidity Reagent RÖTELN VIRUS / RUBELLA VIRUS

Best.-Nr. / Order No.	B129AVID
Chargen-Nr. / Lot No.	SHH.CR
Verfallsdatum / Expiry Date	2027-08
Verwendung mit / For use with SERION ELISA <i>classic</i> Röteln Virus / Rubella Virus IgG	SEH.BD
Verwendung mit Standardserum / For use with standard serum	SEH.AI
Standard Zielwert (SERION Aviditätsindex) / Standard Target Value (SERION Avidity Index)	58%
Standard Zielbereich (SERION Aviditätsindex) / Standard Target Range (SERION Avidity Index)	$\geq 46 \%$ und / and $\leq 70 \%$
X Faktor / X factor	0,347
Y Faktor / Y factor	1,498
Rekonstitution / Reconstitution	2,0 ml Aqua dest.
Versions-Nr. / Version No.	2017 / 09-1



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de