

Analysenzertifikat

Certificate of Analysis

SERION KBR Positivkontrolle / CFT Positive Control PARAINFLUENZA VIRUS 2

Best.-Nr. / Order No.	3117
Chargen-Nr. / Lot No.	SII.CQ
Verfallsdatum / Expiry Date	2028-09
Material / Material	Kaninchenserum / rabbit serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	1:80



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: info@virion-serion.de
 Internet: www.virion-serion.de