



# Analysenzertifikat Certificate of Analysis

virion\sersion

## SERION KBR Positivkontrolle / CFT Positive Control CHLAMYDIA

Best.-Nr. / Order No.	3122
Chargen-Nr. / Lot No.	SBC.EB
Verfallsdatum / Expiry Date	2022-02
Material / Material	Schafsseren / sheep sera
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	1:40

serion



multianalyt™

Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)