



# Analysenzertifikat Certificate of Analysis

virion\serion

## SERION KBR Positivkontrolle / CFT Positive Control CAMPYLOBACTER JEJUNI

Best.-Nr. / Order No.	3206
Chargen-Nr. / Lot No.	SCH.BL
Verfallsdatum / Expiry Date	2027-03
Material / Material	Meerschweinchenserum/guinea pig serum
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	0,1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:10
KBR Titer / CFT Titre	1:80



Hersteller / Manufacturer:  
Institut Virion\Serion GmbH  
Friedrich-Bergius-Ring 19  
D-97076 Würzburg  
Germany  
Tel. +49 931 3045 0  
Fax +49 931 3045 100  
E-Mail: [dialog@virion-serion.de](mailto:dialog@virion-serion.de)  
Internet: [www.virion-serion.de](http://www.virion-serion.de)