



Analysenzertifikat Certificate of Analysis

virion\sersion

SERION KBR Antigen / CFT Antigen ADENOVIRUS

Best.-Nr. / Order No.	1121
Chargen-Nr. / Lot No.	SGB.DW
Verfallsdatum / Expiry Date	2021-07
Antigen / Antigen	Adenoviren / Adenoviruses 2, 3, 4, 7 (Adenoid 6, GB, RI-67, Gomen)
Kultursystem / Culture System	Verozellen / Vero cells
Präparation / Preparation	Mischung aus infizierten Zellen und Zellkulturmedium (konzentriert und gereinigt) Mixture of infected cells and cell culture medium (concentrated and purified)
Inaktivierung / Inactivation	Hitzebehandlung / heat treatment
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:32

serion



multianalyt™

Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de