

Analysenzertifikat Certificate of Analysis

SERION KBR Antigen / CFT Antigen ADENOVIRUS

Best.-Nr. / Order No.	1121
Chargen-Nr. / Lot No.	SKH.DG
Verfallsdatum / Expiry Date	2027-10
Antigen / Antigen	Adenoviren / Adenoviruses 2, 3, 4, 7 (Adenoid 6, GB, RI-67, Gomen)
Kultursystem / Culture System	Verozellen / Vero cells
Präparation / Preparation	Mischung aus infizierten Zellen und Zellkulturmedium (konzentriert und gereinigt) Mixture of infected cells and cell culture medium (concentrated and purified)
Inaktivierung / Inactivation	Hitzebehandlung / heat treatment
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:32



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: dialog@virion-serion.de
 Internet: www.virion-serion.de