



Analysenzertifikat Certificate of Analysis

virion\sersion

SERION KBR Antigen / CFT Antigen CYTOMEGALOVIRUS

Best.-Nr. / Order No.	1130
Chargen-Nr. / Lot No.	SDF.BB
Verfallsdatum / Expiry Date	2025-04
Antigen / Antigen	Cytomegalovirus (AD 169)
Kultursystem / Culture System	HEL 299 Zellen / cells
Präparation / Preparation	infizierte Zellen und Zellkulturmedium (konzentriert und gereinigt) infected cells and culture fluid (concentrated and purified)
Inaktivierung / Inactivation	Hitzebehandlung für 30 min bei 56°C Heat treatment for 30 min at 56°C
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:8



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de

FB-IV-047 B