



Analysenzertifikat Certificate of Analysis

virion\serion

SERION KBR Antigen / CFT Antigen COXSACKIEVIRUS B1

Best.-Nr. / Order No.	1172
Chargen-Nr. / Lot No.	SDH.BP
Verfallsdatum / Expiry Date	2027-04
Antigen / Antigen	Coxsackie B1 (Stamm / strain Conn-5)
Kultursystem / Culture System	Verozellen / Vero cells
Präparation / Preparation	Infizierte Zellen und Zellkulturmedium (konzentriert und gereinigt) / Infected cells and cell culture medium (concentrated and purified)
Inaktivierung / Inactivation	UV-Bestrahlung / UV irradiation
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:4



Hersteller / Manufacturer:
Institut Virion\Serion GmbH
Friedrich-Bergius-Ring 19
D-97076 Würzburg
Germany
Tel. +49 931 3045 0
Fax +49 931 3045 100
E-Mail: dialog@virion-serion.de
Internet: www.virion-serion.de

FB-IV-047 B