

Analysenzertifikat

Certificate of Analysis

SERION KBR Antigen / CFT Antigen COXSACKIEVIRUS B4

Best.-Nr. / Order No.	1175
Chargen-Nr. / Lot No.	SBI.BR
Verfallsdatum / Expiry Date	2028-02
Antigen / Antigen	Coxsackievirus B4 (JVB)
Kultursystem / Culture System	Verozellen / Vero cells
Präparation / Preparation	infizierte Zellen und Zellkulturmedium (konzentriert und gereinigt) infected cells and cell culture medium (concentrated and purified)
Inaktivierung / Inactivation	UV Bestrahlung / UV irradiation
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:4



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: dialog@virion-serion.de
 Internet: www.virion-serion.de