

Analysenzertifikat

Certificate of Analysis

SERION KBR Antigen / CFT Antigen INFLUENZA A/B VIRUS POOL

Best.-Nr. / Order No.	1114
Chargen-Nr. / Lot No.	VDI.AA
Verfallsdatum / Expiry Date	2019-11
Antigen / Antigen	Influenza A Virus (Beijing) + Influenza B Virus (Harbin)
Kultursystem / Culture System	bebrütete Hühnereier/ incubated hen's eggs
Präparation / Preparation	konzentrierte Virus Core-Partikel / concentrated virus core particles
Inaktivierung / Inactivation	Behandlung mit Detergens / treatment with detergent
Konservierung / Preservation	0,01 % Thiomersal / 0.01 % thiomersal
Lagerung / Storage	2 °C - 8 °C
Rekonstitution / Reconstitution	1 ml Aqua dest.
Gebrauchsverdünnung / Working Dilution	1:16



Hersteller / Manufacturer:
 Institut Virion\Serion GmbH
 Friedrich-Bergius-Ring 19
 D-97076 Würzburg
 Germany
 Tel. +49 931 3045 0
 Fax +49 931 3045 100
 E-Mail: dialog@virion-serion.de
 Internet: www.virion-serion.de